

## Infant Radiant Warmer-IRW03

Product Model : **IRW03**



**Category: Baby Care Equipment, Medical Equipment, Infant Radiant Warmer,**

**Legacy of Excellence in Healthcare...**

 [info@gwsmed.com](mailto:info@gwsmed.com), [gwssurgicals@gmail.com](mailto:gwssurgicals@gmail.com)

 +91 124 4915548  [www.gwsmed.com](http://www.gwsmed.com)



**Corp Office:** 1205, 12<sup>th</sup> Floor Ocus Quantum Sector 51, Gurugram - 122003 (Hr.) India

**Works:** Plot No. 38, Sector 7, Sidcul Industrial Area, Haridwar-249403 (Uttarakhand), India

## Description

- Advance micro processor controlled system with self check on the electronic circuitry skin / Air Manual control. Digital display of baby temperature.
- Manual / Servo control modes provide the clinician more flexibility while optimizing the thermal environment.
- In addition the skin / Air temperature display is larger than the other set temperature display.
- Skin temperature can be converted from Celsius to Fahrenheit by the push of a button.
- The spacious baby bed incorporates easy drop down sides for maximum access to the baby.
- The control panel is also the heart for a completely integrated alarm system with both audible and visual indicators in the event of
  - Probe fail
  - Skin High
  - Skin Low
  - Power Fail
  - Heater Fail
- **Manual Mode:**  
Facility of converting the warmer from servo mode to manual mode for non-stop working of the unit.
- **Heater Assembly:**  
The heater assembly located above the centre of the support structure consists of an integral examination light and a 650 watts heating system (Silica or quartz /Ceramic Heater) for quick warming and even distribution of heat on the baby bed. The heater can be swivelled in either direction for taking X-ray and doing procedures.
- **Castors:**  
The system is fitted with 4? high quality castors having brakes in front 2 castors. Antistatic castors  
**(Optional)**
- **Baby Bed:**  
The spacious baby bed incorporates easy drop down/ fold down glasses for maximum access to the patient.  
Optional: X-Ray facility is provided below the baby bed for taking X-ray of the baby without disturbing.
- **Dimensions**

Height	1850mm
Length	1050mm
Width	650mm
- **Electrical Specification (IEC complaint)**

Power supply – 200V / 50Hz

Heater Power – 650 Watts

Fuse – 5 Amp.

- **Highlight of the product**

The unit is designed in such a way that the main control system can be swivelled to the other side for access from all the four sides to the patient in critical condition.

option : weighing scale